


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101818\
Data File : VX005417.D
Acq On : 18 Oct 2018 09:09
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	2706	PASS
75	95	30	60	47.6	7877	PASS
95	95	100	100	100.0	16555	PASS
96	95	5	9	6.9	1135	PASS
173	174	0.00	2	1.3	221	PASS
174	95	50	100	99.7	16511	PASS
175	174	5	9	7.4	1221	PASS
176	174	95	101	97.7	16135	PASS
177	176	5	9	5.7	916	PASS