

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013124.D
 Acq On : 18 Oct 2019 13:49
 Operator : JC/SP
 Sample : VX1018WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1018WBL01

Quant Time: Oct 19 03:44:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	154179	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	241510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	201548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	77104	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	91379	47.42	ug/l	0.00
Spiked Amount						
			Recovery	=	94.84%	
35) Dibromofluoromethane	5.48	113	72109	52.02	ug/l	0.00
Spiked Amount						
			Recovery	=	104.04%	
50) Toluene-d8	8.71	98	280622	51.91	ug/l	0.00
Spiked Amount						
			Recovery	=	103.82%	
62) 4-Bromofluorobenzene	11.13	95	94115	44.80	ug/l	0.00
Spiked Amount						
			Recovery	=	89.60%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013124.D
Acq On : 18 Oct 2019 13:49
Operator : JC/SP
Sample : VX1018WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1018WBL01

Quant Time: Oct 19 03:44:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

