

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013126.D
 Acq On : 18 Oct 2019 14:48
 Operator : JC/SP
 Sample : K5414-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04-SRI4

Quant Time: Oct 19 03:55:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	161031	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	255402	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	216084	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	80917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	95305	47.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.70%	
35) Dibromofluoromethane	5.49	113	75352	51.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.80%	
50) Toluene-d8	8.71	98	294586	51.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.06%	
62) 4-Bromofluorobenzene	11.14	95	98969	44.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.10%	
Target Compounds						
20) Methylene Chloride	2.85	84	2710	1.390	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013126.D
Acq On : 18 Oct 2019 14:48
Operator : JC/SP
Sample : K5414-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04-SRI4

Quant Time: Oct 19 03:55:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

