

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013136.D
 Acq On : 18 Oct 2019 18:41
 Operator : JC/SP
 Sample : K5451-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW7332

Quant Time: Oct 19 05:13:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	162856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	258761	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	222095	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	85722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	100826	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
35) Dibromofluoromethane	5.49	113	78159	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
50) Toluene-d8	8.71	98	298444	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
62) 4-Bromofluorobenzene	11.13	95	101571	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
Target Compounds						
16) Acetone	2.44	43	6487	6.046	ug/l	97
37) Ethyl Acetate	4.83	43	11594	4.558	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013136.D
Acq On : 18 Oct 2019 18:41
Operator : JC/SP
Sample : K5451-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW7332

Quant Time: Oct 19 05:13:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

