

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013139.D
 Acq On : 18 Oct 2019 19:51
 Operator : JC/SP
 Sample : K5451-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JPZ0341

Quant Time: Oct 19 05:29:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	164255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	259003	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	213116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	79993	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	99318	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
35) Dibromofluoromethane	5.49	113	77591	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
50) Toluene-d8	8.71	98	295737	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	11.14	95	94436	41.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.84%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	2508	1.223	ug/l	88
30) Chloroform	5.20	83	1748	0.535	ug/l	94
37) Ethyl Acetate	4.83	43	10024	3.937	ug/l	98
44) Trichloroethene	7.22	130	9813	5.262	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013139.D
Acq On : 18 Oct 2019 19:51
Operator : JC/SP
Sample : K5451-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JPZ0341

Quant Time: Oct 19 05:29:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

