

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013142.D
 Acq On : 18 Oct 2019 21:01
 Operator : JC/SP
 Sample : K5451-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JPZ0348

Quant Time: Oct 21 11:12:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	164567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	265549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	225800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	90961	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	99729	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
35) Dibromofluoromethane	5.48	113	79675	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
50) Toluene-d8	8.71	98	302604	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	11.14	95	104343	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
Target Compounds						
16) Acetone	2.43	43	7693	7.096	ug/l	93
27) cis-1,2-Dichloroethene	4.59	96	9566	4.656	ug/l	89
37) Ethyl Acetate	4.83	43	5587	2.140	ug/l #	95
44) Trichloroethene	7.21	130	40976	21.429	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013142.D
Acq On : 18 Oct 2019 21:01
Operator : JC/SP
Sample : K5451-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JPZ0348

Quant Time: Oct 21 11:12:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

