

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101819\
 Data File : VX013127.D
 Acq On : 18 Oct 2019 15:11
 Operator : JC/SP
 Sample : K5414-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB-01-SRI4

Quant Time: Oct 19 04:16:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	167360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	269481	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	225362	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	89355	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	101225	48.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.78%	
35) Dibromofluoromethane	5.49	113	80147	51.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.64%	
50) Toluene-d8	8.71	98	313878	52.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.13	95	105341	44.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.88%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101819\
Data File : VX013127.D
Acq On : 18 Oct 2019 15:11
Operator : JC/SP
Sample : K5414-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB-01-SRI4

Quant Time: Oct 19 04:16:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.00 min

Lab File: VX013127.D

Acq: 18 Oct 2019 15:11

Instrument :

MSVOA_X

ClientSampleId :


RB-01-SRI4

Tot Ion:168 Resp: 167360

Ion Ratio Lower Upper




168 100

99 59.8 49.4 74.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 48.392 ug/l

RT: 6.05 min Scan# 847

Delta R.T. -0.01 min

Lab File: VX013127.D


Acq: 18 Oct 2019 15:11

Tgt Ion: 65 Resp: 101225

Ion Ratio Lower Upper



65 100


67 51.3 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. 0.00 min

Lab File: VX013127.D

Acq: 18 Oct 2019 15:11

Instrument :

MSVOA_X

ClientSampled :

RB-01-SRI4


Tot Ion:114 Resp: 269481

Ion Ratio Lower Upper

114 100

63 19.9 0.0 40.8




88 17.1 0.0 33.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.819 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX013127.D

Acq: 18 Oct 2019 15:11


Tgt Ion:113 Resp: 80147

Ion Ratio Lower Upper

113 100

111 101.4 83.2 124.8



192 19.9 15.4 23.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.032 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX013127.D
Acq: 18 Oct 2019 15:11




Instrument :
MSVOA_X
ClientSampleId :
RB-01-SRI4


Tot Ion: 98 Resp: 313878
Ion Ratio Lower Upper
98 100
100 66.8 53.4 80.2



#62
4-Bromofluorobenzene
Concen: 44.942 ug/l
RT: 11.13 min Scan# 1680
Delta R.T. -0.01 min
Lab File: VX013127.D
Acq: 18 Oct 2019 15:11

Tgt Ion: 95 Resp: 105341
Ion Ratio Lower Upper
95 100
174 73.5 0.0 143.4
176 69.7 0.0 136.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1512
Delta R.T. -0.01 min
Lab File: VX013127.D
Acq: 18 Oct 2019 15:11

Instrument :
MSVOA_X
ClientSampled :
RB-01-SRI4

Tot Ion:117 Resp: 225362
Ion Ratio Lower Upper
117 100
82 58.4 44.1 66.1
119 31.7 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX013127.D
Acq: 18 Oct 2019 15:11

Tgt Ion:152 Resp: 89355
Ion Ratio Lower Upper
152 100
115 62.2 44.5 133.5
150 155.7 0.0 353.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

