


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101819\
Data File : VX013122.D
Acq On : 18 Oct 2019 11:28
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Title : SW846 8260
Last Update : Wed Oct 16 07:26:22 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	6028	PASS
75	95	30	60	51.2	17359	PASS
95	95	100	100	100.0	33893	PASS
96	95	5	9	7.1	2400	PASS
173	174	0.00	2	0.7	172	PASS
174	95	50	100	71.1	24114	PASS
175	174	5	9	7.9	1897	PASS
176	174	95	101	96.9	23358	PASS
177	176	5	9	6.6	1531	PASS