

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101823\
 Data File : VX038355.D
 Acq On : 18 Oct 2023 12:20
 Operator : JC/MD
 Sample : 04936-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1027

Quant Time: Oct 19 01:38:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132359	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90937	42.715	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.440%
35) Dibromofluoromethane	5.385	113	74168	44.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.500%
50) Toluene-d8	8.647	98	291166	45.464	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.920%#
62) 4-Bromofluorobenzene	11.079	95	113576	46.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
25) 2-Butanone	4.574	43	6369	4.443	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101823\
Data File : VX038355.D
Acq On : 18 Oct 2023 12:20
Operator : JC/MD
Sample : 04936-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1027

Quant Time: Oct 19 01:38:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

