

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101823\
 Data File : VX038363.D
 Acq On : 18 Oct 2023 15:25
 Operator : JC/MD
 Sample : 04940-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-20-23-26

Quant Time: Oct 19 01:41:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205515	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81750	44.056	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.120%
35) Dibromofluoromethane	5.385	113	66031	44.589	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.180%
50) Toluene-d8	8.647	98	260606	46.054	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.100%
62) 4-Bromofluorobenzene	11.079	95	99423	46.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.680%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1256	1.622	ug/l	91
44) Trichloroethene	7.135	130	1706	1.142	ug/l	66
64) Tetrachloroethene	9.275	164	1468	1.077	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101823\
Data File : VX038363.D
Acq On : 18 Oct 2023 15:25
Operator : JC/MD
Sample : 04940-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-20-23-26

Quant Time: Oct 19 01:41:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

