



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101823\
Data File : VX038346.D
Acq On : 18 Oct 2023 08:25
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Title : SW846 8260
Last Update : Thu Oct 05 15:27:13 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.0	9148	PASS
75	95	30	60	55.8	24296	PASS
95	95	100	100	100.0	43520	PASS
96	95	5	9	6.9	2982	PASS
173	174	0.00	2	1.2	431	PASS
174	95	50	100	80.8	35160	PASS
175	174	5	9	7.6	2668	PASS
176	174	95	101	96.7	34008	PASS
177	176	5	9	6.4	2170	PASS