

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101824\
 Data File : VX043430.D
 Acq On : 18 Oct 2024 13:37
 Operator : JC/MD
 Sample : P4427-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 241016065-06-TRIP-BLANK

Quant Time: Oct 18 23:18:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	77159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	142121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51074	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65620	52.074	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.140%
35) Dibromofluoromethane	5.379	113	46401	47.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.680%
50) Toluene-d8	8.646	98	171218	50.409	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.820%
62) 4-Bromofluorobenzene	11.079	95	59583	48.286	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101824\
Data File : VX043430.D
Acq On : 18 Oct 2024 13:37
Operator : JC/MD
Sample : P4427-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
241016065-06-TRIP-BLANK

Quant Time: Oct 18 23:18:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

