

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101824\
 Data File : VX043433.D
 Acq On : 18 Oct 2024 14:46
 Operator : JC/MD
 Sample : P4438-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190-TB-20241015

Quant Time: Oct 18 23:20:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72096	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	130812	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49573	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60217	51.142	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.280%	
35) Dibromofluoromethane	5.379	113	42620	47.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.480%	
50) Toluene-d8	8.647	98	157445	50.361	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.720%	
62) 4-Bromofluorobenzene	11.079	95	57572	50.690	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.380%	
Target Compounds						
16) Acetone	2.386	43	913	1.857	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101824\
Data File : VX043433.D
Acq On : 18 Oct 2024 14:46
Operator : JC/MD
Sample : P4438-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190-TB-20241015

Quant Time: Oct 18 23:20:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

