



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101824\
Data File : VX043423.D
Acq On : 18 Oct 2024 07:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Title : SW846 8260
Last Update : Wed Oct 02 16:50:57 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	4463	PASS
75	95	30	60	55.9	11982	PASS
95	95	100	100	100.0	21435	PASS
96	95	5	9	7.1	1516	PASS
173	174	0.00	2	0.8	119	PASS
174	95	50	100	73.5	15762	PASS
175	174	5	9	7.4	1171	PASS
176	174	95	101	96.8	15258	PASS
177	176	5	9	6.8	1036	PASS