

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101918\
 Data File : VX005452.D
 Acq On : 19 Oct 2018 12:33
 Operator : JC/MD
 Sample : J5195-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JC-02-101818-B

Quant Time: Oct 20 05:00:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	298174	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	436111	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	366625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174805	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160894	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
35) Dibromofluoromethane	5.50	113	133739	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
50) Toluene-d8	8.71	98	480047	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	11.14	95	158887	42.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.30%	
Target Compounds						
					Qvalue	
16) Acetone	2.45	43	14794	7.732	ug/l	97
17) Carbon Disulfide	2.57	76	15309	1.935	ug/l	97
18) Methyl Acetate	2.78	43	5993	1.101	ug/l	99
25) 2-Butanone	4.70	43	5200	1.857	ug/l	95
43) Isopropyl Acetate	6.46	43	180278	24.915	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101918\
Data File : VX005452.D
Acq On : 19 Oct 2018 12:33
Operator : JC/MD
Sample : J5195-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JC-02-101818-B

Quant Time: Oct 20 05:00:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

