

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101918\
 Data File : VX005455.D
 Acq On : 19 Oct 2018 13:53
 Operator : JC/MD
 Sample : J5538-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WWT1-6

Quant Time: Oct 20 05:07:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	381619	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	339857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166024	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	149244	51.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.32%	
35) Dibromofluoromethane	5.49	113	122078	48.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.60%	
50) Toluene-d8	8.72	98	453213	52.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.46%	
62) 4-Bromofluorobenzene	11.14	95	147185	45.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.30%	

Target Compounds

						Qvalue
16) Acetone	2.45	43	39742	23.563	ug/l	99
18) Methyl Acetate	2.78	43	10512	2.191	ug/l	100
20) Methylene Chloride	2.86	84	69426	27.666	ug/l	98
25) 2-Butanone	4.71	43	9826	3.982	ug/l	97
29) Tetrahydrofuran	5.17	42	2603	1.654	ug/l	95
43) Isopropyl Acetate	6.46	43	150526	23.774	ug/l	96
68) m/p-Xylenes	10.36	106	4711	1.033	ug/l	98
95) Naphthalene	13.83	128	451348	39.420	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101918\
Data File : VX005455.D
Acq On : 19 Oct 2018 13:53
Operator : JC/MD
Sample : J5538-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WWT1-6

Quant Time: Oct 20 05:07:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

