

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101918\
 Data File : VX005459.D
 Acq On : 19 Oct 2018 15:40
 Operator : JC/MD
 Sample : J5597-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 22 09:24:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256583	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	362445	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162946	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	151065	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
35) Dibromofluoromethane	5.51	113	122118	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	8.72	98	439908	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
62) 4-Bromofluorobenzene	11.14	95	146094	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101918\
Data File : VX005459.D
Acq On : 19 Oct 2018 15:40
Operator : JC/MD
Sample : J5597-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 22 09:24:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

