

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101918\
 Data File : VX005458.D
 Acq On : 19 Oct 2018 15:13
 Operator : JC/MD
 Sample : J5597-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 22 09:24:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	374852	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	334628	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	163924	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151654	51.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.24%	
35) Dibromofluoromethane	5.49	113	126313	50.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.76%	
50) Toluene-d8	8.71	98	443126	51.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.98%	
62) 4-Bromofluorobenzene	11.14	95	146123	45.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.28%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101918\
Data File : VX005458.D
Acq On : 19 Oct 2018 15:13
Operator : JC/MD
Sample : J5597-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 22 09:24:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX005458.D

Acq: 19 Oct 2018 15:13

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-5

Tot Ion:168 Resp: 267299

Ion Ratio Lower Upper

168 100

99 49.7 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 51.120 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.01 min

Lab File: VX005458.D


Acq: 19 Oct 2018 15:13

Tgt Ion: 65 Resp: 151654

Ion Ratio Lower Upper

65 100

67 52.3 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20 6.30









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.86 min Scan# 936
Delta R.T. -0.01 min
Lab File: VX005458.D
Acq: 19 Oct 2018 15:13

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tot Ion:114 Resp: 374852
Ion Ratio Lower Upper
114 100
63 20.2 0.0 39.0
88 19.1 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VX0
Ion 87.90 (87.60 to 88.60): VX0






#35
Dibromofluoromethane
Concen: 50.876 ug/l
RT: 5.49 min Scan# 712
Delta R.T. -0.01 min
Lab File: VX005458.D
Acq: 19 Oct 2018 15:13

Tgt Ion:113 Resp: 126313
Ion Ratio Lower Upper
113 100
111 101.8 82.4 123.6
192 23.0 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.989 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX005458.D

Acq: 19 Oct 2018 15:13

Instrument :

MSVOA_X

ClientSampleId :





STORAGE-BLANK-WATER-REF-5

Tot Ion: 98 Resp: 443126

Ion Ratio Lower Upper

98 100

100 63.6 52.9 79.3



#62

4-Bromofluorobenzene

Concen: 45.636 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.00 min

Lab File: VX005458.D

Acq: 19 Oct 2018 15:13




Tgt Ion: 95 Resp: 146123


Ion Ratio Lower Upper

95 100

174 96.4 0.0 185.2

176 92.4 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX005458.D
Acq: 19 Oct 2018 15:13

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tot Ion:117 Resp: 334628
Ion Ratio Lower Upper
117 100
82 53.3 47.8 71.6
119 32.4 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX005458.D
Acq: 19 Oct 2018 15:13

Tgt Ion:152 Resp: 163924
Ion Ratio Lower Upper
152 100
115 54.6 39.0 117.0
150 154.0 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

