


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101918\
Data File : VX005444.D
Acq On : 19 Oct 2018 07:49
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	7399	PASS
75	95	30	60	48.0	20451	PASS
95	95	100	100	100.0	42650	PASS
96	95	5	9	6.8	2908	PASS
173	174	0.00	2	1.4	539	PASS
174	95	50	100	91.5	39029	PASS
175	174	5	9	7.9	3078	PASS
176	174	95	101	97.6	38077	PASS
177	176	5	9	6.7	2558	PASS