

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101920\
 Data File : VX018975.D
 Acq On : 20 Oct 2020 00:48
 Operator : JC/SP
 Sample : L4413-01 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30688

Quant Time: Oct 20 09:58:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 19 18:41:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	91389	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	175675	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	158127	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	71532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	65380	56.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.00%	
35) Dibromofluoromethane	5.47	113	54091	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
50) Toluene-d8	8.70	98	219563	52.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.12%	
62) 4-Bromofluorobenzene	11.12	95	75062	50.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.38%	
Target Compounds						
64) Tetrachloroethene	9.32	164	11823	10.582	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101920\
Data File : VX018975.D
Acq On : 20 Oct 2020 00:48
Operator : JC/SP
Sample : L4413-01 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30688

Quant Time: Oct 20 09:58:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101920W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 19 18:41:11 2020
Response via : Initial Calibration

