

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101921\
 Data File : VX024849.D
 Acq On : 19 Oct 2021 15:28
 Operator : JC/MD
 Sample : M4213-01ME
 Misc : 3.48g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JDGE3ME

Quant Time: Oct 20 04:23:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML100721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 20 04:21:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	261082	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	235403	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	117583	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	57690	35.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.38%
7) Chloroethane-d5	1.648	69	35694	33.54	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.08%#
11) 1,1-Dichloroethene-d2	2.282	63	91920	28.96	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.92%#
21) 2-Butanone-d5	4.489	46	154562	116.30	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.30%
24) Chloroform-d	5.068	84	141331	43.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.34%
26) 1,2-Dichloroethane-d4	5.964	65	101356	43.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.54%
32) Benzene-d6	5.976	84	271696	40.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.94%
36) 1,2-Dichloropropane-d6	7.312	67	95739	46.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.32%
41) Toluene-d8	8.653	98	249268	38.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.16%#
43) trans-1,3-Dichloroprop...	8.952	79	41456	40.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.68%
47) 2-Hexanone-d5	9.403	63	101994	104.75	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.75%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	133643	46.79	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.58%
66) 1,2-Dichlorobenzene-d4	12.323	152	105197	45.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.46%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101921\
Data File : VX024849.D
Acq On : 19 Oct 2021 15:28
Operator : JC/MD
Sample : M4213-01ME
Misc : 3.48g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JDGE3ME

Quant Time: Oct 20 04:23:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM100721WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 20 04:21:44 2021
Response via : Initial Calibration

