

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032206.D
 Acq On : 19 Oct 2022 14:31
 Operator : JC/MD
 Sample : VX1019MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1019MBL01

Quant Time: Oct 19 16:56:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	88120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	122128	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125650	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80082	69.858	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 139.720%#	
35) Dibromofluoromethane	5.385	113	65629	77.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 154.440%#	
50) Toluene-d8	8.647	98	245660	80.302	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 160.600%#	
62) 4-Bromofluorobenzene	11.079	95	82905	74.007	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 148.020%#	
Target Compounds						
17) Carbon Disulfide	2.514	76	1899	1.031	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032206.D
Acq On : 19 Oct 2022 14:31
Operator : JC/MD
Sample : VX1019MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1019MBL01

Quant Time: Oct 19 16:56:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

