

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032213.D
 Acq On : 19 Oct 2022 17:23
 Operator : JC/MD
 Sample : N5114-01
 Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31266

Quant Time: Oct 19 19:55:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	176612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	276060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.030	152	138828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132360	57.609	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 115.220%		
35) Dibromofluoromethane	5.391	113	99174	51.626	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.260%		
50) Toluene-d8	8.647	98	385325	55.723	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 111.440%		
62) 4-Bromofluorobenzene	11.085	95	151398	59.790	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 119.580%		
Target Compounds						
					Qvalue	
25) 2-Butanone	4.599	43	12933	11.110	ug/l	92
39) Methylcyclohexane	7.379	83	13965	4.731	ug/l	94
52) Toluene	8.720	92	103002	22.201	ug/l	100
67) Ethyl Benzene	10.195	91	70967	8.136	ug/l	98
68) m/p-Xylenes	10.299	106	106988	31.059	ug/l	98
69) o-Xylene	10.640	106	43898	13.303	ug/l	99
73) Isopropylbenzene	10.963	105	21091	2.129	ug/l	98
78) n-propylbenzene	11.305	91	79391	6.999	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	84755	9.905	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	343838	41.333	ug/l	100
85) sec-Butylbenzene	11.896	105	55033	5.319	ug/l	88
86) p-Isopropyltoluene	12.012	119	31908	3.625	ug/l	99
89) n-Butylbenzene	12.335	91	130453	16.824	ug/l #	55

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032213.D
Acq On : 19 Oct 2022 17:23
Operator : JC/MD
Sample : N5114-01
Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31266

Quant Time: Oct 19 19:55:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

