

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032214.D
 Acq On : 19 Oct 2022 17:46
 Operator : JC/MD
 Sample : N5114-03 20X
 Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FMC-0080

Quant Time: Oct 19 19:51:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	162377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219618	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	111800	52.927	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.860%			
35) Dibromofluoromethane	5.385	113	85255	48.124	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.240%			
50) Toluene-d8	8.647	98	314582	49.330	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.660%			
62) 4-Bromofluorobenzene	11.079	95	111614	47.797	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.600%			
Target Compounds						
39) Methylcyclohexane	7.379	83	4483	1.647	ug/l	95
52) Toluene	8.720	92	241120	56.356	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032214.D
Acq On : 19 Oct 2022 17:46
Operator : JC/MD
Sample : N5114-03 20X
Misc : 1.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FMC-0080

Quant Time: Oct 19 19:51:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

