

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032231.D
 Acq On : 20 Oct 2022 01:05
 Operator : JC/MD
 Sample : N5164-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 588A-2

Quant Time: Oct 20 05:23:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224799	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94417	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95004	52.151	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.300%
35) Dibromofluoromethane	5.385	113	72466	46.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.640%
50) Toluene-d8	8.647	98	265342	47.122	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.240%
62) 4-Bromofluorobenzene	11.079	95	96049	46.581	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032231.D
Acq On : 20 Oct 2022 01:05
Operator : JC/MD
Sample : N5164-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
588A-2

Quant Time: Oct 20 05:23:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

