

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032233.D
 Acq On : 20 Oct 2022 01:51
 Operator : JC/MD
 Sample : N5166-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5388-2

Quant Time: Oct 20 05:24:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222271	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194491	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90934	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97448	54.628	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.260%
35) Dibromofluoromethane	5.391	113	75041	48.517	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	8.647	98	273374	49.100	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	98933	48.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032233.D
Acq On : 20 Oct 2022 01:51
Operator : JC/MD
Sample : N5166-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5388-2

Quant Time: Oct 20 05:24:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

