

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032234.D
 Acq On : 20 Oct 2022 02:14
 Operator : JC/MD
 Sample : N5176-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5236-1

Quant Time: Oct 20 05:24:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197253	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90174	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98422	55.042	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.080%
35) Dibromofluoromethane	5.385	113	75392	48.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.480%
50) Toluene-d8	8.647	98	271770	48.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.600%
62) 4-Bromofluorobenzene	11.079	95	98080	47.607	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032234.D
Acq On : 20 Oct 2022 02:14
Operator : JC/MD
Sample : N5176-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5236-1

Quant Time: Oct 20 05:24:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

