

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032241.D
 Acq On : 20 Oct 2022 04:55
 Operator : JC/MD
 Sample : PB148361ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#07

Quant Time: Oct 20 06:02:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91331	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96749	54.963	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.920%
35) Dibromofluoromethane	5.391	113	73602	47.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.560%
50) Toluene-d8	8.647	98	267248	47.696	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.400%
62) 4-Bromofluorobenzene	11.079	95	96218	46.895	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.780%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13852	24.249	ug/l	100
20) Methylene Chloride	2.788	84	6764	4.612	ug/l	89
37) Ethyl Acetate	4.721	43	17738	9.048	ug/l	99
43) Isopropyl Acetate	6.342	43	87521	28.444	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032241.D
Acq On : 20 Oct 2022 04:55
Operator : JC/MD
Sample : PB148361ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#07

Quant Time: Oct 20 06:02:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

