

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032244.D
 Acq On : 20 Oct 2022 06:05
 Operator : JC/MD
 Sample : PB148361ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#10

Quant Time: Oct 20 13:12:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99312	62.212	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 124.420%#	
35) Dibromofluoromethane	5.385	113	74460	51.313	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.620%	
50) Toluene-d8	8.647	98	253008	48.436	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.880%	
62) 4-Bromofluorobenzene	11.079	95	92018	48.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.220%	
Target Compounds						
16) Acetone	2.386	43	13490	26.040	ug/l	96
20) Methylene Chloride	2.794	84	6484	4.874	ug/l	96
37) Ethyl Acetate	4.715	43	18726	10.246	ug/l	99
43) Isopropyl Acetate	6.342	43	88737	30.935	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032244.D
Acq On : 20 Oct 2022 06:05
Operator : JC/MD
Sample : PB148361ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#10

Quant Time: Oct 20 13:12:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

