

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
 Data File : VX032245.D
 Acq On : 20 Oct 2022 06:28
 Operator : JC/MD
 Sample : PB148361ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#11

Quant Time: Oct 20 13:13:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	101948	60.646	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	121.300%#
35) Dibromofluoromethane	5.391	113	75239	49.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.180%
50) Toluene-d8	8.653	98	255328	46.279	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.560%
62) 4-Bromofluorobenzene	11.079	95	93877	46.467	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13824	25.340	ug/l	99
20) Methylene Chloride	2.788	84	6836	4.880	ug/l	95
37) Ethyl Acetate	4.715	43	18293	9.477	ug/l	99
43) Isopropyl Acetate	6.336	43	88315	29.150	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101922\
Data File : VX032245.D
Acq On : 20 Oct 2022 06:28
Operator : JC/MD
Sample : PB148361ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#11

Quant Time: Oct 20 13:13:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

