

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
 Data File : VX038381.D
 Acq On : 19 Oct 2023 13:28
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 20 03:39:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209880	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116963	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83281	44.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.980%
35) Dibromofluoromethane	5.385	113	67092	44.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.720%
50) Toluene-d8	8.647	98	265524	45.947	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	102121	46.612	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
Data File : VX038381.D
Acq On : 19 Oct 2023 13:28
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 20 03:39:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

