

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101923\
 Data File : VX038370.D
 Acq On : 19 Oct 2023 08:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:34:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	52.198	-4.4	95	0.00
3 P	Chloromethane	50.000	47.899	4.2	97	0.00
4 C	Vinyl Chloride	50.000	53.354	-6.7#	100	0.00
5 T	Bromomethane	50.000	54.863	-9.7	95	0.00
6 T	Chloroethane	50.000	54.391	-8.8	94	0.00
7 T	Trichlorofluoromethane	50.000	54.786	-9.6	102	0.00
8 T	Diethyl Ether	50.000	51.328	-2.7	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.579	-7.2	101	0.00
10 T	Methyl Iodide	50.000	46.914	6.2	88	0.00
11 T	Tert butyl alcohol	250.000	212.538	15.0	81	0.00
12 CM	1,1-Dichloroethene	50.000	51.577	-3.2#	96	0.00
13 T	Acrolein	250.000	146.544	41.4#	61	0.00
14 T	Allyl chloride	50.000	46.849	6.3	88	0.00
15 T	Acrylonitrile	250.000	235.408	5.8	88	0.00
16 T	Acetone	250.000	289.592	-15.8	113	0.00
17 T	Carbon Disulfide	50.000	45.683	8.6	90	0.00
18 T	Methyl Acetate	50.000	48.068	3.9	91	0.00
19 T	Methyl tert-butyl Ether	50.000	48.457	3.1	91	0.00
20 T	Methylene Chloride	50.000	44.166	11.7	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.042	-0.1	92	0.00
22 T	Diisopropyl ether	50.000	50.592	-1.2	92	0.00
23 T	Vinyl Acetate	250.000	253.582	-1.4	92	0.00
24 P	1,1-Dichloroethane	50.000	48.211	3.6	91	0.00
25 T	2-Butanone	250.000	249.183	0.3	94	-0.01
26 T	2,2-Dichloropropane	50.000	52.630	-5.3	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.023	-2.0	97	0.00
28 T	Bromochloromethane	50.000	47.033	5.9	84	0.00
29 T	Tetrahydrofuran	250.000	233.819	6.5	89	0.00
30 C	Chloroform	50.000	52.604	-5.2#	96	0.00
31 T	Cyclohexane	50.000	51.627	-3.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	52.342	-4.7	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.034	5.9	87	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	91	0.00
35 S	Dibromofluoromethane	50.000	50.704	-1.4	92	0.00
36 T	1,1-Dichloropropene	50.000	51.182	-2.4	96	0.00
37 T	Ethyl Acetate	50.000	48.691	2.6	90	-0.01
38 T	Carbon Tetrachloride	50.000	53.450	-6.9	96	0.00
39 T	Methylcyclohexane	50.000	53.950	-7.9	97	0.00
40 TM	Benzene	50.000	52.576	-5.2	95	0.00
41 T	Methacrylonitrile	50.000	50.674	-1.3	92	-0.01
42 TM	1,2-Dichloroethane	50.000	50.525	-1.0	92	0.00
43 T	Isopropyl Acetate	50.000	50.326	-0.7	89	0.00
44 TM	Trichloroethene	50.000	55.451	-10.9	100	0.00
45 C	1,2-Dichloropropane	50.000	52.946	-5.9#	96	0.00
46 T	Dibromomethane	50.000	53.289	-6.6	98	0.00
47 T	Bromodichloromethane	50.000	56.699	-13.4	98	0.00
48 T	Methyl methacrylate	50.000	51.882	-3.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101923\
 Data File : VX038370.D
 Acq On : 19 Oct 2023 08:58
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:34:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	864.870	13.5	77	0.00
50 S	Toluene-d8	50.000	50.127	-0.3	91	0.00
51 T	4-Methyl-2-Pentanone	250.000	264.889	-6.0	96	0.00
52 CM	Toluene	50.000	54.649	-9.3#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	57.064	-14.1	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	56.679	-13.4	98	0.00
55 T	1,1,2-Trichloroethane	50.000	54.985	-10.0	99	0.00
56 T	Ethyl methacrylate	50.000	54.187	-8.4	95	0.00
57 T	1,3-Dichloropropane	50.000	54.569	-9.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	237.345	5.1	85	0.00
59 T	2-Hexanone	250.000	273.391	-9.4	99	0.00
60 T	Dibromochloromethane	50.000	58.944	-17.9	101	0.00
61 T	1,2-Dibromoethane	50.000	55.746	-11.5	99	0.00
62 S	4-Bromofluorobenzene	50.000	52.813	-5.6	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	55.123	-10.2	103	0.00
65 PM	Chlorobenzene	50.000	52.910	-5.8	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	57.707	-15.4	103	0.00
67 C	Ethyl Benzene	50.000	53.502	-7.0#	100	0.00
68 T	m/p-Xylenes	100.000	109.862	-9.9	103	0.00
69 T	o-Xylene	50.000	53.341	-6.7	101	0.00
70 T	Styrene	50.000	55.914	-11.8	103	0.00
71 P	Bromoform	50.000	51.659	-3.3	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	52.008	-4.0	104	0.00
74 T	N-ethyl acetate	50.000	51.156	-2.3	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.459	-2.9	103	0.00
76 T	1,2,3-Trichloropropane	50.000	51.847	-3.7	104	0.00
77 T	Bromobenzene	50.000	50.991	-2.0	104	0.00
78 T	n-propylbenzene	50.000	51.624	-3.2	103	0.00
79 T	2-Chlorotoluene	50.000	49.201	1.6	102	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.553	-5.1	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.258	7.5	96	0.00
82 T	4-Chlorotoluene	50.000	50.377	-0.8	102	0.00
83 T	tert-Butylbenzene	50.000	52.799	-5.6	106	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.199	-4.4	103	0.00
85 T	sec-Butylbenzene	50.000	52.589	-5.2	104	0.00
86 T	p-Isopropyltoluene	50.000	53.531	-7.1	106	0.00
87 T	1,3-Dichlorobenzene	50.000	51.405	-2.8	105	0.00
88 T	1,4-Dichlorobenzene	50.000	51.352	-2.7	108	0.00
89 T	n-Butylbenzene	50.000	52.959	-5.9	105	0.00
90 T	Hexachloroethane	50.000	55.618	-11.2	106	0.00
91 T	1,2-Dichlorobenzene	50.000	53.295	-6.6	110	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.497	-1.0	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.625	-7.2	111	0.00
94 T	Hexachlorobutadiene	50.000	57.508	-15.0	119	0.00
95 T	Naphthalene	50.000	51.107	-2.2	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
Data File : VX038370.D
Acq On : 19 Oct 2023 08:58
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 20 03:34:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.501	-7.0	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6