

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
 Data File : VX038389.D
 Acq On : 19 Oct 2023 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:42:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.483	0.548	-13.5	96	0.00
3 P	Chloromethane	0.492	0.506	-2.8	97	0.00
4 C	Vinyl Chloride	0.545	0.610	-11.9#	97	0.00
5 T	Bromomethane	0.744	0.887	-19.2	100	0.00
6 T	Chloroethane	0.484	0.442	8.7	88	0.00
7 T	Trichlorofluoromethane	1.077	1.222	-13.5	98	0.00
8 T	Diethyl Ether	0.402	0.419	-4.2	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.535	-13.8	99	0.00
10 T	Methyl Iodide	0.677	0.678	-0.1	87	0.00
11 T	Tert butyl alcohol	0.196	0.160	18.4	79	0.00
12 CM	1,1-Dichloroethene	0.457	0.514	-12.5#	97	0.00
13 T	Acrolein	0.114	0.092	19.3	78	0.00
14 T	Allyl chloride	0.818	0.815	0.4	87	0.00
15 T	Acrylonitrile	0.366	0.371	-1.4	88	0.00
16 T	Acetone	0.336	0.308	8.3	83	0.00
17 T	Carbon Disulfide	1.318	1.238	6.1	86	0.00
18 T	Methyl Acetate	1.439	1.537	-6.8	94	0.00
19 T	Methyl tert-butyl Ether	1.822	1.884	-3.4	90	0.00
20 T	Methylene Chloride	0.631	0.609	3.5	92	0.00
21 T	trans-1,2-Dichloroethene	0.566	0.600	-6.0	91	0.00
22 T	Diisopropyl ether	1.715	1.781	-3.8	88	0.00
23 T	Vinyl Acetate	1.258	1.302	-3.5	87	0.00
24 P	1,1-Dichloroethane	1.041	1.047	-0.6	88	0.00
25 T	2-Butanone	0.542	0.532	1.8	86	0.00
26 T	2,2-Dichloropropane	0.813	0.814	-0.1	85	0.00
27 T	cis-1,2-Dichloroethene	0.675	0.721	-6.8	94	0.00
28 T	Bromochloromethane	0.470	0.446	5.1	78	0.00
29 T	Tetrahydrofuran	0.353	0.350	0.8	88	0.00
30 C	Chloroform	1.138	1.235	-8.5#	92	0.00
31 T	Cyclohexane	0.871	0.880	-1.0	85	0.00
32 T	1,1,1-Trichloroethane	0.975	1.040	-6.7	89	0.00
33 S	1,2-Dichloroethane-d4	0.804	0.743	7.6	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.360	0.373	-3.6	87	0.00
36 T	1,1-Dichloropropene	0.497	0.524	-5.4	91	0.00
37 T	Ethyl Acetate	0.590	0.596	-1.0	86	0.00
38 T	Carbon Tetrachloride	0.489	0.541	-10.6	91	0.00
39 T	Methylcyclohexane	0.582	0.646	-11.0	92	0.00
40 TM	Benzene	1.390	1.502	-8.1	90	0.00
41 T	Methacrylonitrile	0.294	0.310	-5.4	88	-0.01
42 TM	1,2-Dichloroethane	0.555	0.583	-5.0	88	0.00
43 T	Isopropyl Acetate	0.881	0.928	-5.3	86	0.00
44 TM	Trichloroethene	0.363	0.422	-16.3	97	0.00
45 C	1,2-Dichloropropane	0.360	0.394	-9.4#	92	0.00
46 T	Dibromomethane	0.279	0.310	-11.1	94	0.00
47 T	Bromodichloromethane	0.496	0.571	-15.1	91	0.00
48 T	Methyl methacrylate	0.420	0.455	-8.3	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
 Data File : VX038389.D
 Acq On : 19 Oct 2023 17:10
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 03:42:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	75	0.00
50 S	Toluene-d8	1.377	1.409	-2.3	86	0.00
51 T	4-Methyl-2-Pentanone	0.608	0.671	-10.4	92	0.00
52 CM	Toluene	0.905	1.026	-13.4#	94	0.00
53 T	t-1,3-Dichloropropene	0.533	0.605	-13.5	89	0.00
54 T	cis-1,3-Dichloropropene	0.560	0.638	-13.9	91	0.00
55 T	1,1,2-Trichloroethane	0.387	0.444	-14.7	96	0.00
56 T	Ethyl methacrylate	0.594	0.663	-11.6	90	0.00
57 T	1,3-Dichloropropane	0.626	0.711	-13.6	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.291	0.298	-2.4	85	0.00
59 T	2-Hexanone	0.488	0.540	-10.7	92	0.00
60 T	Dibromochloromethane	0.379	0.466	-23.0	98	0.00
61 T	1,2-Dibromoethane	0.419	0.491	-17.2	96	0.00
62 S	4-Bromofluorobenzene	0.522	0.573	-9.8	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.324	0.372	-14.8	100	0.00
65 PM	Chlorobenzene	1.046	1.143	-9.3	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.415	-17.2	98	0.00
67 C	Ethyl Benzene	1.868	2.070	-10.8#	96	0.00
68 T	m/p-Xylenes	0.722	0.802	-11.1	97	0.00
69 T	o-Xylene	0.705	0.771	-9.4	97	0.00
70 T	Styrene	1.148	1.314	-14.5	98	0.00
71 P	Bromoform	0.292	0.340	-16.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.353	3.580	-6.8	99	0.00
74 T	N-amyl acetate	1.432	1.495	-4.4	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.285	-4.5	97	0.00
76 T	1,2,3-Trichloropropane	1.137	1.190	-4.7	97	0.00
77 T	Bromobenzene	0.819	0.889	-8.5	102	0.00
78 T	n-propylbenzene	4.100	4.327	-5.5	98	0.00
79 T	2-Chlorotoluene	2.432	2.477	-1.9	97	0.00
80 T	1,3,5-Trimethylbenzene	2.893	3.096	-7.0	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.356	0.347	2.5	86	0.00
82 T	4-Chlorotoluene	2.877	2.929	-1.8	95	0.00
83 T	tert-Butylbenzene	2.827	3.016	-6.7	98	0.00
84 T	1,2,4-Trimethylbenzene	2.884	3.076	-6.7	97	0.00
85 T	sec-Butylbenzene	3.662	3.937	-7.5	98	0.00
86 T	p-Isopropyltoluene	3.021	3.318	-9.8	100	0.00
87 T	1,3-Dichlorobenzene	1.644	1.725	-4.9	99	0.00
88 T	1,4-Dichlorobenzene	1.679	1.740	-3.6	100	0.00
89 T	n-Butylbenzene	3.009	3.146	-4.6	96	0.00
90 T	Hexachloroethane	0.501	0.550	-9.8	96	0.00
91 T	1,2-Dichlorobenzene	1.581	1.694	-7.1	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.297	0.301	-1.3	90	0.00
93 T	1,2,4-Trichlorobenzene	1.060	1.144	-7.9	103	0.00
94 T	Hexachlorobutadiene	0.427	0.470	-10.1	105	0.00
95 T	Naphthalene	3.692	3.779	-2.4	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101923\
Data File : VX038389.D
Acq On : 19 Oct 2023 17:10
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 20 03:42:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.026	1.087	-5.9	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6