


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102020\
Data File : VX018977.D
Acq On : 20 Oct 2020 15:22
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102020W.M
Title : SW846 8260
Last Update : Tue Oct 20 19:16:32 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	8360	PASS
75	95	30	60	51.1	23763	PASS
95	95	100	100	100.0	46538	PASS
96	95	5	9	6.4	2982	PASS
173	174	0.00	2	0.9	337	PASS
174	95	50	100	78.0	36317	PASS
175	174	5	9	7.3	2666	PASS
176	174	95	101	95.1	34549	PASS
177	176	5	9	6.6	2276	PASS