

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
 Data File : VX032264.D
 Acq On : 20 Oct 2022 15:43
 Operator : JC/MD
 Sample : N5210-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MID1022

Quant Time: Oct 21 01:26:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94357	48.129	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.260%
35) Dibromofluoromethane	5.391	113	72754	45.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	271843	47.244	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	93259	44.261	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
Data File : VX032264.D
Acq On : 20 Oct 2022 15:43
Operator : JC/MD
Sample : N5210-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MID1022

Quant Time: Oct 21 01:26:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

