

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
 Data File : VX032270.D
 Acq On : 20 Oct 2022 18:00
 Operator : JC/MD
 Sample : N5185-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-02

Quant Time: Oct 21 02:15:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238628	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98639	48.789	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.391	113	74303	44.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.500%
50) Toluene-d8	8.647	98	282722	47.298	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.600%
62) 4-Bromofluorobenzene	11.079	95	99666	45.534	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.060%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.111	73	12156	2.543	ug/l #	9
40) Benzene	6.043	78	9421	1.522	ug/l	91
67) Ethyl Benzene	10.195	91	26263	3.474	ug/l	99
73) Isopropylbenzene	10.963	105	2934	0.421	ug/l	99
78) n-propylbenzene	11.305	91	3794	0.476	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
Data File : VX032270.D
Acq On : 20 Oct 2022 18:00
Operator : JC/MD
Sample : N5185-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-02

Quant Time: Oct 21 02:15:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration

