

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
 Data File : VX032254.D
 Acq On : 20 Oct 2022 11:47
 Operator : JC/MD
 Sample : VX1020WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1020WBL01

Quant Time: Oct 21 01:21:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 19 16:51:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	252089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213199	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99901	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	105424	49.550	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.100%
35) Dibromofluoromethane	5.385	113	79026	45.049	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.100%
50) Toluene-d8	8.647	98	304036	48.148	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.300%
62) 4-Bromofluorobenzene	11.079	95	107150	46.339	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.680%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102022\
Data File : VX032254.D
Acq On : 20 Oct 2022 11:47
Operator : JC/MD
Sample : VX1020WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01

Quant Time: Oct 21 01:21:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 19 16:51:09 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01


Tgt Ion:168 Resp: 163552
Ion Ratio Lower Upper
168 100
99 58.8 44.8 67.2



#33
1,2-Dichloroethane-d4
Concen: 49.550 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47

Tgt Ion: 65 Resp: 105424
Ion Ratio Lower Upper
65 100
67 49.9 0.0 103.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01


Tgt Ion:114 Resp: 252089
Ion Ratio Lower Upper
114 100
63 20.4 0.0 38.6
88 15.8 0.0 32.6



#35
Dibromofluoromethane
Concen: 45.049 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47

Tgt Ion:113 Resp: 79026
Ion Ratio Lower Upper
113 100
111 102.8 82.8 124.2
192 19.5 15.5 23.3









#50
Toluene-d8
Concen: 48.148 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01


Tgt Ion: 98 Resp: 304036
Ion Ratio Lower Upper
98 100
100 63.1 50.7 76.1



#62
4-Bromofluorobenzene
Concen: 46.339 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47

Tgt Ion: 95 Resp: 107150
Ion Ratio Lower Upper
95 100
174 78.7 0.0 166.8
176 75.8 0.0 161.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47

Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01

Tgt Ion:117 Resp: 213199
Ion Ratio Lower Upper
117 100
82 56.8 44.2 66.4
119 33.6 25.8 38.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX032254.D
Acq: 20 Oct 2022 11:47

Tgt Ion:152 Resp: 99901
Ion Ratio Lower Upper
152 100
115 60.5 43.1 129.4
150 155.0 0.0 348.6

