

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
 Data File : VX038393.D
 Acq On : 20 Oct 2023 10:29
 Operator : JC/MD
 Sample : VX1020WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1020WBL01

Quant Time: Oct 21 01:26:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122122	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	221058	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115621	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87860	44.729	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.460%
35) Dibromofluoromethane	5.385	113	71729	45.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.647	98	273819	44.986	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.980%#
62) 4-Bromofluorobenzene	11.079	95	102598	44.462	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
Data File : VX038393.D
Acq On : 20 Oct 2023 10:29
Operator : JC/MD
Sample : VX1020WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1020WBL01

Quant Time: Oct 21 01:26:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

