

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
 Data File : VX038398.D
 Acq On : 20 Oct 2023 13:34
 Operator : JC/MD
 Sample : 04985-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 21 01:30:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226663	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90762	44.695	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.380%
35) Dibromofluoromethane	5.385	113	73873	45.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.460%
50) Toluene-d8	8.647	98	282446	45.256	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.520%#
62) 4-Bromofluorobenzene	11.079	95	106110	44.847	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
Data File : VX038398.D
Acq On : 20 Oct 2023 13:34
Operator : JC/MD
Sample : 04985-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Oct 21 01:30:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

