

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
 Data File : VX038399.D
 Acq On : 20 Oct 2023 13:57
 Operator : JC/MD
 Sample : 04986-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-101923

Quant Time: Oct 21 01:39:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224444	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90652	45.188	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.380%
35) Dibromofluoromethane	5.385	113	73595	45.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	283256	45.835	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.660%#
62) 4-Bromofluorobenzene	11.079	95	104031	44.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
Data File : VX038399.D
Acq On : 20 Oct 2023 13:57
Operator : JC/MD
Sample : 04986-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-101923

Quant Time: Oct 21 01:39:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

