

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
 Data File : VX038402.D
 Acq On : 20 Oct 2023 15:06
 Operator : JC/MD
 Sample : 04986-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-06-209-229-101923

Quant Time: Oct 21 01:42:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208747	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83896	44.503	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.000%
35) Dibromofluoromethane	5.385	113	69129	45.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.647	98	264090	45.947	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	102785	47.170	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.340%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	3305	4.201	ug/l	99
48) Methyl methacrylate	7.696	41	4604	2.623	ug/l	92
52) Toluene	8.720	92	132373	35.028	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
Data File : VX038402.D
Acq On : 20 Oct 2023 15:06
Operator : JC/MD
Sample : 04986-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-06-209-229-101923

Quant Time: Oct 21 01:42:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

