

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
 Data File : VX038399.D
 Acq On : 20 Oct 2023 13:57
 Operator : JC/MD
 Sample : 04986-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-101923

Quant Time: Oct 21 01:39:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224444	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229683	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90652	45.188	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.380%
35) Dibromofluoromethane	5.385	113	73595	45.506	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.020%
50) Toluene-d8	8.647	98	283256	45.835	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.660%#
62) 4-Bromofluorobenzene	11.079	95	104031	44.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.800%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102023\
Data File : VX038399.D
Acq On : 20 Oct 2023 13:57
Operator : JC/MD
Sample : 04986-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-101923

Quant Time: Oct 21 01:39:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57




Instrument :
MSVOA_X
ClientSampleId :
TB-01-101923


Tgt Ion:168 Resp: 124724
Ion Ratio Lower Upper
168 100
99 58.1 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 45.188 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57

Tgt Ion: 65 Resp: 90652
Ion Ratio Lower Upper
65 100
67 50.9 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57




Instrument :
MSVOA_X
ClientSampleId :
TB-01-101923


Tgt Ion:114 Resp: 224444
Ion Ratio Lower Upper
114 100
63 19.7 0.0 45.0
88 15.5 0.0 36.6



#35
Dibromofluoromethane
Concen: 45.506 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57

Tgt Ion:113 Resp: 73595
Ion Ratio Lower Upper
113 100
111 101.8 82.9 124.3
192 18.1 13.3 19.9









#50
Toluene-d8
Concen: 45.835 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57




Instrument :
MSVOA_X
ClientSampleId :
TB-01-101923


Tgt Ion: 98 Resp: 283256
Ion Ratio Lower Upper
98 100
100 62.1 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 44.403 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038399.D
Acq: 20 Oct 2023 13:57

Tgt Ion: 95 Resp: 104031
Ion Ratio Lower Upper
95 100
174 83.1 0.0 135.6
176 78.0 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX038399.D

Acq: 20 Oct 2023 13:57

Instrument :

MSVOA_X

ClientSampleId :

TB-01-101923





Tgt Ion:117 Resp: 229683

Ion Ratio Lower Upper

117 100

82 54.3 46.2 69.4

119 32.5 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX038399.D

Acq: 20 Oct 2023 13:57

Tgt Ion:152 Resp: 116798

Ion Ratio Lower Upper

152 100

115 60.1 43.3 129.9

150 156.2 0.0 345.0

