

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
 Data File : VX048241.D
 Acq On : 20 Oct 2025 16:18
 Operator : JC/MD
 Sample : Q3371-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB184-HYD-20251010

Quant Time: Oct 20 16:48:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	127530	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.745	114	218641	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	203489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	103207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	99453	55.621	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.240%
35) Dibromofluoromethane	5.367	113	78492	52.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.180%
50) Toluene-d8	8.635	98	66280	12.668	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	25.340%#
62) 4-Bromofluorobenzene	11.061	95	101289	50.890	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.780%
Target Compounds						
						Qvalue
60) Dibromochloromethane	9.506	129	1096	0.560	ug/l	100
85) sec-Butylbenzene	11.872	105	1674	0.217	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
Data File : VX048241.D
Acq On : 20 Oct 2025 16:18
Operator : JC/MD
Sample : Q3371-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB184-HYD-20251010

Quant Time: Oct 20 16:48:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

