

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
 Data File : VX048245.D
 Acq On : 20 Oct 2025 17:41
 Operator : JC/MD
 Sample : Q3371-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-915-917

Quant Time: Oct 20 18:52:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	107423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	183824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	169870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	81048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	76776	50.975	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.960%
35) Dibromofluoromethane	5.379	113	60761	48.417	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.635	98	207133	47.089	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.180%
62) 4-Bromofluorobenzene	11.061	95	75303	45.000	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.000%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	33048	63.960	ug/l	98
17) Carbon Disulfide	2.520	76	1264	0.356	ug/l	97
25) 2-Butanone	4.550	43	8182	12.511	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
Data File : VX048245.D
Acq On : 20 Oct 2025 17:41
Operator : JC/MD
Sample : Q3371-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-915-917

Quant Time: Oct 20 18:52:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

