

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
 Data File : VX048246.D
 Acq On : 20 Oct 2025 18:02
 Operator : JC/MD
 Sample : Q3371-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-920-922

Quant Time: Oct 20 18:54:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	122446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	210478	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	193818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	92863	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95954	55.892	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.780%
35) Dibromofluoromethane	5.373	113	76505	53.242	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.480%
50) Toluene-d8	8.635	98	258724	51.369	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.740%
62) 4-Bromofluorobenzene	11.061	95	93348	48.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102025\
Data File : VX048246.D
Acq On : 20 Oct 2025 18:02
Operator : JC/MD
Sample : Q3371-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-920-922

Quant Time: Oct 20 18:54:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

