

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102119\
 Data File : VX013162.D
 Acq On : 21 Oct 2019 19:14
 Operator : JC/SP
 Sample : K5441-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-1

Manual Integrations
 APPROVED

MMDadoda
 10/22/2019 12:32:28 PM

Quant Time: Oct 22 02:27:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	26380	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	150833	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	130340	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.05	65	62410	30.05	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	100.17%	
60) 4-Bromofluorobenzene	11.14	95	63633	28.04	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	93.47%	
63) Toluene-d8	8.71	98	179732	31.00	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	103.33%	
Target Compounds						
15) Acetone	2.43	58	4476	15.545	ug/l	Ovalue 96
30) 2-Butanone	4.67	43	3307m	2.691	ug/l	
58) 4-Methyl-2-Pentanone	8.64	43	7438	3.387	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102119\
Data File : VX013162.D
Acq On : 21 Oct 2019 19:14
Operator : JC/SP
Sample : K5441-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-1

Manual Integrations
APPROVED

MMDadoda
10/22/2019 12:32:28 PM

Quant Time: Oct 22 02:27:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 22 01:53:07 2019
Response via : Initial Calibration

