

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102120\
 Data File : VX019012.D
 Acq On : 21 Oct 2020 20:37
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX102120

Quant Time: Oct 22 18:28:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.498	0.499	-0.2	98	0.00
3 P	Chloromethane	0.495	0.486	1.8	100	0.00
4 C	Vinyl Chloride	0.584	0.580	0.7#	98	0.00
5 T	Bromomethane	0.596	0.601	-0.8	110	0.00
6 T	Chloroethane	0.365	0.427	-17.0	107	0.00
7 T	Trichlorofluoromethane	0.987	0.954	3.3	97	0.00
8 T	Diethyl Ether	0.387	0.376	2.8	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.452	3.4	100	0.00
10 T	Methyl Iodide	0.590	0.609	-3.2	96	0.00
11 T	Tert butyl alcohol	0.121	0.119	1.7	102	0.00
12 CM	1,1-Dichloroethene	0.487	0.474	2.7#	101	0.00
13 T	Acrolein	0.048	0.045	6.3	107	0.00
14 T	Allyl chloride	0.630	0.620	1.6	101	0.00
15 T	Acrylonitrile	0.235	0.234	0.4	101	0.00
16 T	Acetone	0.204	0.203	0.5	104	0.00
17 T	Carbon Disulfide	1.404	1.332	5.1	100	0.00
18 T	Methyl Acetate	0.506	0.507	-0.2	101	0.00
19 T	Methyl tert-butyl Ether	1.542	1.516	1.7	99	0.00
20 T	Methylene Chloride	0.568	0.551	3.0	102	0.00
21 T	trans-1,2-Dichloroethene	0.568	0.552	2.8	101	0.00
22 T	Diisopropyl ether	1.241	1.224	1.4	99	0.00
23 T	Vinyl Acetate	1.179	1.188	-0.8	101	0.00
24 P	1,1-Dichloroethane	0.904	0.892	1.3	100	0.00
25 T	2-Butanone	0.322	0.330	-2.5	102	0.00
26 T	2,2-Dichloropropane	0.835	0.786	5.9	96	0.00
27 T	cis-1,2-Dichloroethene	0.644	0.641	0.5	102	0.00
28 T	Bromochloromethane	0.384	0.370	3.6	99	0.00
29 T	Tetrahydrofuran	0.193	0.191	1.0	101	0.00
30 C	Chloroform	0.982	0.991	-0.9#	101	0.00
31 T	Cyclohexane	0.701	0.665	5.1	96	0.00
32 T	1,1,1-Trichloroethane	0.890	0.885	0.6	99	0.00
33 S	1,2-Dichloroethane-d4	0.599	0.575	4.0	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.317	0.304	4.1	102	0.00
36 T	1,1-Dichloropropene	0.449	0.438	2.4	100	0.00
37 T	Ethyl Acetate	0.383	0.388	-1.3	101	0.00
38 T	Carbon Tetrachloride	0.471	0.456	3.2	97	0.00
39 T	Methylcyclohexane	0.503	0.476	5.4	96	0.00
40 TM	Benzene	1.321	1.317	0.3	100	0.00
41 T	Methacrylonitrile	0.206	0.203	1.5	101	0.00
42 TM	1,2-Dichloroethane	0.451	0.452	-0.2	100	0.00
43 T	Isopropyl Acetate	0.649	0.651	-0.3	100	0.00
44 TM	Trichloroethene	0.385	0.376	2.3	99	0.00
45 C	1,2-Dichloropropane	0.317	0.317	0.0	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\ VX102120\
 Data File : VX019012.D
 Acq On : 21 Oct 2020 20:37
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX102120

Quant Time: Oct 22 18:28:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.241	0.241	0.0	102	0.00
47 T	Bromodichloromethane	0.473	0.478	-1.1	100	0.00
48 T	Methyl methacrylate	0.320	0.313	2.2	101	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	102	0.00
50 S	Toluene-d8	1.208	1.168	3.3	102	0.00
51 T	4-Methyl-2-Pentanone	0.382	0.399	-4.5	101	0.00
52 CM	Toluene	0.873	0.873	0.0	100	0.00
53 T	t-1,3-Dichloropropene	0.534	0.543	-1.7	102	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.571	-1.2	101	0.00
55 T	1,1,2-Trichloroethane	0.356	0.355	0.3	100	0.00
56 T	Ethyl methacrylate	0.512	0.526	-2.7	100	0.00
57 T	1,3-Dichloropropane	0.565	0.577	-2.1	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.269	0.272	-1.1	102	0.00
59 T	2-Hexanone	0.285	0.304	-6.7	103	0.00
60 T	Dibromochloromethane	0.388	0.405	-4.4	101	0.00
61 T	1,2-Dibromoethane	0.381	0.383	-0.5	100	0.00
62 S	4-Bromofluorobenzene	0.430	0.429	0.2	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.381	0.351	7.9	100	0.00
65 PM	Chlorobenzene	1.049	1.027	2.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.373	0.0	99	0.00
67 C	Ethyl Benzene	1.808	1.742	3.7#	99	0.00
68 T	m/p-Xylenes	0.693	0.679	2.0	99	0.00
69 T	o-Xylene	0.663	0.655	1.2	100	0.00
70 T	Styrene	1.125	1.143	-1.6	102	0.00
71 P	Bromoform	0.301	0.312	-3.7	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.597	3.451	4.1	99	0.00
74 T	N-amyl acetate	1.152	1.174	-1.9	103	0.00
75 P	1,1,2,2-Tetrachloroethane	1.084	1.075	0.8	102	0.00
76 T	1,2,3-Trichloropropane	0.999	1.011	-1.2	105	0.00
77 T	Bromobenzene	0.908	0.878	3.3	101	0.00
78 T	n-propylbenzene	4.076	3.908	4.1	99	0.00
79 T	2-Chlorotoluene	2.378	2.287	3.8	100	0.00
80 T	1,3,5-Trimethylbenzene	2.970	2.895	2.5	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.390	0.5	100	0.00
82 T	4-Chlorotoluene	2.829	2.726	3.6	100	0.00
83 T	tert-Butylbenzene	2.926	2.916	0.3	101	0.00
84 T	1,2,4-Trimethylbenzene	2.974	2.914	2.0	100	0.00
85 T	sec-Butylbenzene	3.541	3.427	3.2	99	0.00
86 T	p-Isopropyltoluene	3.235	3.181	1.7	100	0.00
87 T	1,3-Dichlorobenzene	1.666	1.561	6.3	101	0.00
88 T	1,4-Dichlorobenzene	1.710	1.640	4.1	104	0.00
89 T	n-Butylbenzene	2.961	2.796	5.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102120\
Data File : VX019012.D
Acq On : 21 Oct 2020 20:37
Operator : JC/SP
Sample : VSTDICV050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX102120

Quant Time: Oct 22 18:28:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.495	0.510	-3.0	101	0.00
91 T	1,2-Dichlorobenzene	1.554	1.533	1.4	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.247	0.240	2.8	103	0.00
93 T	1,2,4-Trichlorobenzene	1.137	1.068	6.1	101	0.00
94 T	Hexachlorobutadiene	0.527	0.491	6.8	104	0.00
95 T	Naphthalene	3.576	3.594	-0.5	103	0.00
96 T	1,2,3-Trichlorobenzene	1.106	1.023	7.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 4