


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102120\
 Data File : VX019005.D
 Acq On : 21 Oct 2020 17:10
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Title : SW846 8260
 Last Update : Thu Oct 22 17:51:42 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	12598	PASS
75	95	30	60	49.2	38915	PASS
95	95	100	100	100.0	79104	PASS
96	95	5	9	6.7	5319	PASS
173	174	0.00	2	0.9	608	PASS
174	95	50	100	81.1	64165	PASS
175	174	5	9	7.6	4871	PASS
176	174	95	101	99.0	63554	PASS
177	176	5	9	6.6	4176	PASS