

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032294.D
 Acq On : 21 Oct 2022 17:19
 Operator : JC/MD
 Sample : N5248-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-TB-20221018

Quant Time: Oct 22 05:12:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150212	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228306	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94882	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94681	48.452	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.900%
35) Dibromofluoromethane	5.391	113	72477	45.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.240%
50) Toluene-d8	8.653	98	270098	47.229	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.460%
62) 4-Bromofluorobenzene	11.079	95	97488	46.552	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	885	1.396	ug/l	90
20) Methylene Chloride	2.788	84	1120	0.688	ug/l #	86
67) Ethyl Benzene	10.195	91	2269	0.311	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	1433	0.252	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032294.D
Acq On : 21 Oct 2022 17:19
Operator : JC/MD
Sample : N5248-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-TB-20221018

Quant Time: Oct 22 05:12:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

