

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032295.D
 Acq On : 21 Oct 2022 17:42
 Operator : JC/MD
 Sample : N5248-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-RB-20221018

Quant Time: Oct 22 05:13:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	143970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92837	49.569	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.140%
35) Dibromofluoromethane	5.391	113	70082	44.684	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.360%
50) Toluene-d8	8.653	98	262464	46.489	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.980%
62) 4-Bromofluorobenzene	11.079	95	93274	45.117	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.240%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1544	2.540	ug/l	94
40) Benzene	6.044	78	2938	0.502	ug/l	94
52) Toluene	8.720	92	11860	3.131	ug/l	98
67) Ethyl Benzene	10.195	91	3162	0.447	ug/l	99
68) m/p-Xylenes	10.305	106	3368	1.206	ug/l	89
69) o-Xylene	10.646	106	1354	0.506	ug/l	91
84) 1,2,4-Trimethylbenzene	11.750	105	2039	0.374	ug/l	96
95) Naphthalene	13.780	128	1669	0.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032295.D
Acq On : 21 Oct 2022 17:42
Operator : JC/MD
Sample : N5248-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-RB-20221018

Quant Time: Oct 22 05:13:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

